# Bruksanvisning – Catell Armsling Standard

Catell rekommenderar att produkten provas ut och anpassas av legitimerad eller övrigt utbildad sjukvårdskunnig personal. Brukare med diabetes, känslig hud, kända kontaktallergier, nedsatt känsel eller dålig cirkulation ska vara extra observant för tryck, skav eller ödem under användning. Kontrollera alltid materialinnehåll nedan. Alla allvarliga incidenter i samband med användning av denna produkt ska anmälas till tillverkaren och till behörig myndighet i den medlemsstat där användaren och/eller patienten är etablerad. Om frågor kring produkten uppstår kontakta din ordinatör, utprovare eller Catell.

**Användningsområde**
Ortosen stödjer och avlastar brukarens arm och skuldra efter skada eller operation.

**Applicering**
1. Öppna upp armpåsens kardborrband och placera underarmen i påsen. Applicera kardborrbandet med plastöglor i handledsnivå. Se till att delen med plastöglan hamnar vänd in mot kroppen. Stäng bandet. *Bild 1*
2. Fäst och stäng det breda korta kardborrbandet över påsens öppning i nivå med armvecket. *Bild 2*
3. För det fasta bandet från armpåsen snett över ryggen och över motsatt axel. Vänd bandet genom den övre plastöglan, fäst med kardborren på samma band. Var noga med att justera in rätt längd för optimalt stöd. *Bild 3*
4. Vid behov av extra stöd för handen, placera det korta, smala bandet i tumvecket mellan tumme och pekfinger och fäst i påsen. *Bild 4a*
5. Vid behov av ytterligare fixering av armen mot bålen används extrabandet som midjebälte. Fäst detta på armpåsen vid armbågen, dra runt bålen och fäst igenom plastöglan vid handledsnivå. *Bild 4a, 4b*
6. Säkerställ att samtliga kardborrband är ordentligt stängda.

Ortosen kan behöva justeras med jämna mellanrum beroende på vilka situationer den används i. Det är viktigt att ortosen sitter bekvämt, utan att klämma, skava eller orsaka stas.

**Material**
Nylon, polyester. Produkten är latexfri.

**Tvättråd**
Slut kardborrbanden. Följ tvättråden nedan. Använd gärna tvättpåse. Låt ortosen dropptorka.

**Måttabell**

| STRL | MÅTT * | STRL | MÅTT * | STRL | MÅTT * | STRL | MÅTT * | STRL | MÅTT * |
|---|---|---|---|---|---|---|---|---|---|
| XS | 23-28 cm | S | 28-33 cm | M | 33-38 cm | L | 38-43 cm | XL | 43-48 cm |

*Mät underarmens längd. Mått ska ses som en vägledning. Produkten är bilateral.
Denna produkt är endast avsedd att användas på en patient.

**Återvinning**
Separera plast och textil. Sortera enligt lokala föreskrifter.

Denna produkt överensstämmer med de krav som ställs i den Medicintekniska förordningen (EU) 2017/745, klass I produkter.

Manufacturer Catell AB
Örkroken 23, SE-138 40 Älta, Sweden | +46 (0)8 727 95 30 | info@catell.se | www.catell.se

# Instruction for use – Catell Armsling Standard

Catell recommend that the product is fitted and adapted by a licensed or otherwise qualified health professionals.Users with diabetes, sensitive skin, known contact allergies, impaired sensibility or poor circulation must be extra observant and check for pinching, chafing or edema when using the brace. Always check material content below. Any serious incidents occurring relates to the device should be reported to the manufacturer and to the competent authority of the Member State in which the user and/or patient is resident. Contact your prescriber or Catell if questions about the product occur.

**Area of use**

The brace supports the users arm and shoulder after injury or surgery.

**Application**

1. Open the arm pouches hook and loop closures and place the forearm in the pouch. Place the strap with plastic buckles at wrist level. Make sure part with plastic buckle end up facing towards body. Close the strap. *Picture 1*
2. Place and close the wide short hook strap over the opening of the pouch, near the bend of the arm. *Picture 2*
3. Pull the strap fixed on the pouch behind your back, over the opposite shoulder, turn through the plastic buckle and attach the hook on the same strap. Thoroughly adjust the correct length of strap for optimal support. *Picture 3*
4. In need of additional support for your hand, place the short narrow strap in the thumb webbing, between thumb and index finger, and attach to the arm pouch. *Picture 4a*
5. In need of extra stability of the arm towards body, the accompanying band is used as a waist belt. Attach the belt to arm pouch at elbow side, place it around your back and turn trough plastic buckle at wrist level. *Picture 4a, 4b*
6. Make sure that all straps are properly secured.

When necessary, adjust the brace. It is important that the brace fits comfortably without pinching, chafing or affecting blood circulation

**Material**

Nylon, polyester. Free from latex.

**Washing instructions**

Close the hook and loop straps. Follow the washing instructions below. Use a washing pouch. Let the brace drip dry.

**Size**

| SIZE | DIMENSIONS * | SIZE | DIMENSIONS * | SIZE | DIMENSIONS * | SIZE | DIMENSIONS * | SIZE | DIMENSIONS * |
|---|---|---|---|---|---|---|---|---|---|
| XS | 23-28 cm | S | 28-33 cm | M | 33-38 cm | L | 38-43 cm | XL | 43-48 cm |

*Measure length of forearm. Measurements should be seen as guidelines. The product is bilateral. Single patient use.

**Recycling**

Separate plastic and textile. Sort according to local regulations.

This device conforms to the requirements of the Medical Device Regulation (EU) 2017/745 product class I.

Manufacturer Catell AB
Örkroken 23, SE-138 40 Älta, Sweden | +46 (0)8 727 95 30 | info@catell.se | www.catell.se