Bruksanvisning – Catell Armsling Smart

Catell rekommenderar att produkten provas ut och anpassas av legitimerad eller övrigt utbildad sjukvårdskunnig personal. Brukare med diabetes, känslig hud, kända kontaktallergier, nedsatt känsel eller dålig cirkulation ska vara extra observant för tryck, skav eller ödem under användning. Kontrollera alltid materialinnehåll nedan. Alla allvarliga incidenter i samband med användning av denna produkt ska anmälas till tillverkaren och till behörig myndighet i den medlemsstat där användaren och/eller patienten är etablerad. Om frågor kring produkten uppstår kontakta din ordinatör, utprovare eller Catell.

Användningsområde

Ortosen stödjer, avlastar och stabiliserar brukarens arm och skuldra efter skada eller operation.

Applicering

Omfånget på bålbandet och överarmsmanschetten kan lätt anpassas med sax. Följ de tvärgående sömmarna och klipp av överblivet material i anslutning till någon av dessa.

1. Öppna upp bålbandet och placera det strax under bröstet, bandet bör gå ca 20 cm omlott. Stäng bålbandet. Tänk på att vända bandet så att kardborrknäppningen lätt kan öppnas med den friska armen/handen. *Bild 1*
2. Placera överarmsmanschetten på överarmen (skadade sidan). Vänd manschetten så att kardborrknäppningen lätt kan öppnas med den friska armen/handen. Förslut manschetten och fäst den på bålbandet. *Bild 2*
3. Placera handledsmanschetten på bålbandet, i nivå med handleden, så att plastsöljan hamnar uppåt. Därefter försluts manschetten.
4. Placera axelremmens kardborrdel på mitten av bålbandets baksida. För axelremmen snett över ryggen och upp på den friska sidans axel.
5. Fäst axelbandet i handledsmanschetten genom att stänga klickspännet. Justera därefter längden på axelremmen. Längden på axelremmen kan lätt anpassas med sax. Justera axelpolstringen så att den hamnar över axeln. *Bild 3*
6. Var noga med positioneringen av skuldra, arm och hand.
7. Vid behov av extra stöd för handen, placera det smala bandet i tumvecket mellan tumme och pekfinger och fäst på handledsmanschetten. *Bild 3*
8. Säkerställ att samtliga kardborrband är ordentligt stängda.

Ortosen kan behöva justeras med jämna mellanrum beroende på vilka situationer den används i. Det är viktigt att ortosen sitter bekvämt, utan att klämma, skava eller orsaka stas.

Material

Nylon, polyester. Produkten är latexfri.

Tvättråd

Slut kardborrbanden. Följ tvättråden nedan. Använd gärna tvättpåse. Låt ortosen dropptorka.

Måttabell

One size. Produkten är bilateral. Denna produkt är endast avsedd att användas på en patient.

Återvinning

Separera plast och textil. Sortera enligt lokala föreskrifter.

Denna produkt överensstämmer med de krav som ställs i den Medicintekniska förordningen (EU) 2017/745, klass I produkter.

CE

Manufacturer Catell AB
Örkroken 23, SE-138 40 Älta, Sweden | +46 (0)8 727 95 30 | info@catell.se | www.catell.se

Instruction for use – Catell Armsling Smart

Catell recommend that the product is fitted and adapted by a licensed or otherwise qualified health professionals.Users with diabetes, sensitive skin, known contact allergies, impaired sensibility or poor circulation must be extra observant and check for pinching, chafing or edema when using the brace. Always check material content below. Any serious incidents occurring relates to the device should be reported to the manufacturer and to the competent authority of the Member State in which the user and/or patient is resident. Contact your prescriber or Catell if questions about the product occur.

Area of use

The brace support and stabilizes the users arm and shoulder after injury or surgery.

Application

The range of the body strap and of the upper arm cuff can easily be adjusted with scissors. Follow the longitudinal seams and cut off any remaining material in connection with any of these.

1. Open the body strap and place it just below the chest. It should have about 20 cm overlap. Close the strap. Remember to turn the strap so that it easily can be opened with the healthy arm / hand. *Picture 1*
2. Place the upper arm cuff on the upper arm (injured side). Turn the opening so that it easily can be opened with the healthy arm / hand. Close the upper arm cuff and attach it to the body strap. *Picture 2*
3. Place the wrist cuff on the body strap, at level of the wrist. Make sure the plastic buckle is facing upwards. Then close the cuff.
4. Place the hook of the shoulder strap on center of the back side of the body strap. Place the shoulder strap obliquely over your back and up towards the shoulder on the healthy side.
5. Attach the shoulder strap to the wrist cuff using click buckle. Adjust the length of the shoulder strap. The length of the shoulder strap can easily be adjusted with scissors. Adjust the shoulder pad so that it lies on the shoulder. *Picture 3*
6. Position the shoulder, arm and hand with care.
7. In need of extra support for the hand, place the short narrow strap in thumb webbing, between thumb and index finger, and attach to wrist cuff. *Picture 3*
8. Make sure that all straps are properly secured.

When necessary adjust the hook and loop straps. It is important that the brace fits comfortably without pinching, chafing or affecting blood circulation.

Material

Nylon, polyester. Free from latex.

Washing instructions

Close the hook and loop straps. Follow the washing instructions below. Use a washing pouch. Let the brace drip dry.

Size

One size. The product is bilateral. Single patient use.

Recycling

Separate plastic and textile. Sort according to local regulations.

This device conforms to the requirements of the Medical Device Regulation (EU) 2017/745 product class I.

Manufacturer Catell AB
Örkroken 23, SE-138 40 Älta, Sweden | +46 (0)8 727 95 30 | info@catell.se | www.catell.se